

**INDIAN INSTITUTE OF TECHNOLOGY  
KHARAGPUR – 721 302, INDIA**

**No. IIT/KGP/S&P/CRF/XPS-UPS/2017-18; Date: 16-11-2017**

*Sub: Procurement of X-ray Photoelectron Spectrometer (XPS) attached with Ultraviolet Photoelectron Spectrometer (UPS) and related accessories*

Indian Institute of Technology Kharagpur invites sealed bids from reputed manufacturer or their authorised Distributor/Dealer (foreign firms or their Indian Representative company) who have adequate credential for supply/installation of X-ray Photoelectron Spectrometer (XPS) attached with ultraviolet photoelectron spectrometer (UPS) with accessories as per Annexure – I for the Central Research Facility, Indian Institute of Technology Kharagpur, India. **The XPS and UPS will facilitate the research activities at the undergraduate, postgraduate and doctoral level, in addition to training of students at the UG and PG laboratory courses and research collaboration/consultancies with the Institutes outside IIT Kharagpur.**

*Interested parties may submit their sealed bids under Two-Bid system as per **Technical Specifications** given at **Annexure I**, and **General Terms & Conditions and Special instructions** given at **Annexure II and Annexure III**. Bidders are required to submit their **Details** in the format given at **Annexure IV** along with their technical bids. They are also required to submit a signed **Declaration** in format given at **Annexure V**. A **Checklist and Compliance statement** against each item of Annexure I-III thereon severally and individually shall also be submitted. Kindly refer to the Institute website [www.iitkgp.ac.in](http://www.iitkgp.ac.in) [link: Tenders] for complete tender details.*

*The tender be sent in a sealed packet, containing two separate sealed envelopes (one each for **Technical Bid and Price Bid**) along with prescribed tender fee and Earnest Money Deposit (EMD) duly superscripted with Tender Notice No IIT/KGP/S&P/CA-XPS/2017-18 **Date: 10-11-2017**, to the Office of the **Central Research Facility, Indian Institute of Technology Kharagpur, P.O. Kharagpur Technology, PIN 721302, West Bengal, India on or before 04-01-2018 by 11.30 a.m.***

*For any query pertaining to this bid document, communication be address to **Central Research Facility, Indian Institute of Technology Kharagpur-721302 [Ph: 913222-282480]**.*

*The technical bid will be evaluated first and price bids will be opened in respect of those OEMs/Vendors, who are found technically qualified after evaluation of Technical bids.*

*The Technical bids will be opened on 04.01.2018 at 3.30 p.m. in the office of Chairman, Central Research Facility, **Indian Institute of Technology Kharagpur-721302.***

*Chairman, Central Research Facility  
FOR DIRECTOR IIT, KHARAGPUR*

*Copy to: 1. Notice Board, 2. Institute website, 3. Central Public Procurement Portal, 4. Establishment for advertisement in selected newspapers*

Indian Institute of Technology, Kharagpur  
Central Research Facility

**Specifications for X-ray Photoelectron Spectrometer (XPS) attached with ultraviolet photoelectron spectrometer (UPS)**

The Central Research Facility wishes to procure a *X-ray Photoelectron Spectrometer (XPS) attached with ultraviolet photoelectron spectrometer (UPS)* with options for studying a wide variety of materials, which could be metallic, ceramic, semiconductor, polymer, rubber, emulsion/gel (frozen condition) and biological samples. The desired system should have monochromatic X-ray source capable enough providing survey and narrow spectra along with elemental and chemical state mapping of **all kinds of materials as stated above**. It must also be attached with UV source to provide valence band spectra of all above samples. The system should have all the necessary components, hardware, and software to meet the required specifications as given below.

The detailed specifications are given below:

A. General	A1. The required XPS system must be able to characterize a broad range of materials including conducting, magnetic, semiconducting, insulating solid materials in the form of bulk or powder or thin film, and soft materials such as organics, rubber, polymers, emulsion/gel (frozen condition) and biological samples. The system should provide survey spectra, region spectra, and mapping (elemental and chemical states imaging) of surface and also allow depth profiling and angle-dependent XPS measurements.
B. Vacuum System	B1. The entry lock chamber should be evacuated by turbomolecular pump (TMP) backed by oil-free scroll pump. B2. The analysis chamber should be evacuated by titanium sublimation pump (TSP), TMP, and oil-free scroll pumps. These pumps must provide a guaranteed pressure $\leq 5 \times 10^{-10}$ torr without a sample in the analysis chamber. B3. There must be suitable pressure read out gauges for both the chambers. The system should have safety features that protect the vacuum in case of power failure. B4. All the pumps and valves must be software controlled.
C. Load-lock chamber	C1. The system must have an entry-lock/load-lock chamber to insert the sample in the first step, i.e., prior to allowing sample into analysis chamber. C2. The load lock chamber must have a software-controlled arrangement to cool or heat the samples in the range of $-100\text{ }^{\circ}\text{C}$ to $700\text{ }^{\circ}\text{C}$ or wider and suitable display unit to read the temperature. In case, entry lock chamber do not have heating/cooling facility, a sample preparation chamber must be included which have heating and cooling provision in the above mentioned temperature range and that must be attached with TMP backed by oil-free scroll pumps. There should be

	<p>suitable mechanism to transfer the sample from sample preparation chamber to analysis chamber without breaking the vacuum.</p> <p>C3. The entry lock chamber must have suitable camera to view the sample holder.</p>
D. Analysis chamber	<p>D1. The analysis chamber must have bakeable fully UHV Mu-metal chamber. The magnetic component inside the analysis chamber should be shielded so that analysis is not affected by the magnetic field.</p> <p>D2. The chamber should have sufficient number of ports to accommodate monochromatic X-ray source, achromatic dual anode source, UV Lamp for UPS, ion sources, electron neutralization, etc.</p> <p>D3. The analysis chamber must have suitable X-ray protected lead-glass windows to view the samples through naked eye.</p> <p>D4. This chamber must have optical camera and/or microscope to view the exact analysis position on the samples.</p>
E. Sample Stage manipulator and Sample holder	<p>E1. The stage manipulator should hold sample holder of size <math>\geq 50</math> mm dia or an area <math>\geq 2400</math> mm<sup>2</sup>. The stage should allow holding up to 12 mm thick/height sample.</p> <p>E2. The stage must have software-controlled 5-axes movements in X, Y, Z, tilt, and continuous rotation.</p> <p>E3. The stage should allow software controlled angle-dependent XPS measurement (0 to 80 degree or wider) at a resolution of <math>\leq 0.5^\circ</math>. While angle being changed, the region of interest must remain same.</p> <p>E4. The stage must have arrangement for resistive heating of samples (up to 700°C or more) and cooling (up to <math>-100</math> °C or lower) with liquid N<sub>2</sub>.</p> <p>E5. Sample holders (of size mentioned in E1) for powder, thin film, and thick sample must be provided (at least two for each type of sample).</p> <p>E6. Holder for angle-dependent measurement.</p> <p>E7. Separate holder suitable for heating and cooling studies.</p>
F. X-ray source	<p>F1. The system must have monochromatic Al K<math>\alpha</math> X-ray source.</p> <p>F2. The analysis area must be in the range of <math>\leq 20</math> <math>\mu</math>m to <math>\geq 300</math> <math>\mu</math>m.</p>
G. Analyzer and Detectors	<p>G1. The electron energy analyzer must have <math>\geq 125</math> mm mean radius with full 180 degree hemispherical analyzer.</p> <p>G2. The detector must have at least 128 channels in both spectroscopy and mapping analysis modes.</p> <p>G3. The energy range of analysis must be 10 – 1100 eV or wider. Minimum step size should be specified at different pass energies.</p>
H. Resolution (spectroscopy and imaging mode)	<p>H1. The XPS energy resolution must be <math>\leq 0.5</math> eV (FWHM) for Ag 3d<sub>5/2</sub> peak and <math>\leq 0.85</math> eV for C 1s of O–C–O of polyester sample.</p> <p>H2. The system should provide elemental and chemical maps with a</p>

	resolution $\leq 20 \mu\text{m}$ .
I. Sensitivity	<p>The minimum sensitivity must be demonstrated as follows:</p> <p>I1. <math>\geq 5</math> kcps on <math>\geq 20 \mu\text{m}</math> analysis area at a resolution of <math>\leq 0.6</math> eV (for FWHM Ag 3d<sub>5/2</sub>)</p> <p>I2. <math>\geq 2</math> Mcps on large analysis area at a resolution of <math>\leq 1.0</math> eV (for FWHM Ag 3d<sub>5/2</sub>)</p>
J. Charge neutralization	<p>J1. Robust charge neutralization mechanism, especially for non-conducting samples, using suitable ion/electron beam.</p> <p>J2. The charge neutralization must work in depth profiling mode too.</p> <p>J3. The capability of charge neutralization must be demonstrated on polyethylene terephthalate (PET) with FWHM &lt; 0.85 eV for O=CO peak in C 1s spectrum.</p>
K. Ion gun and Depth profiling	<p>K1. To study the depth profiling of a wide variety of samples, a high performance ion gun capable of operating in both monoatomic and cluster ion modes must be provided. Also provide the available cluster sizes (or number of Ar atoms per cluster) that can be obtained in the cluster mode.</p> <p>K2. The energy of monoatomic Ar<sup>+</sup> ion source should be in the range of 500eV to 4 keV or wider.</p> <p>K3. For the Ar<sup>+</sup> ion cluster, the source energy should be <math>\geq 15</math> keV. Provide the etch rate with standard SiO<sub>2</sub> sample.</p>
L. UPS system	<p>L1. The system must be provided with a high intensity UPS source (UV lamp) producing He I / He II for valence band spectroscopy and work function measurement.</p> <p>L2. The sensitivity of UPS system should be <math>\geq 1</math> Mcps for Ag 4d and the Fermi edge <math>\leq 120</math> meV in width at room temperature. Resolution here is defined as the energy difference between the points at which the intensity is 20% and 80% of the maximum value below the Fermi level.</p>
M. REELS	M1 To detect and quantify hydrogen on the surface of materials.
N. Other accessories (Chiller, compressor, UPS etc.)	<p>N1. Branded chiller and compressor with adequate capacity (as required for your system) must be supplied to support the quoted system for smooth operation. Brand should be mentioned.</p> <p>N2. Branded On-line UPS with battery ( kVA rating based on power requirement of the instrument) for uninterrupted power supply for at least 1 h backup.</p> <p>N3. Necessary number of high purity N<sub>2</sub>, Ar, and He gas cylinders, as required for installation and operation of the instrument.</p> <p>N4. Stainless-steel gas lines</p>

O. PC/Computer	<p>O1. The current generation PC (Intel Core i5/7 with 62 bit latest window 7 or 10 operating system, Intel Xeon or equivalent processor, <math>\geq 8</math> Gb RAM, <math>\geq 1</math> TB hard disk) which supports the software and smooth operation of XPS, UPS, and other attached systems must be included. The supplier should upgrade the software whenever they are available.</p> <p>O2. The PC should have a monitor with <math>\geq 24</math>" TFT/LED screen size along with keyboard and mouse.</p> <p>O3. The PC should have CD/DVD writer, Ethernet and USB ports.</p>
P. Software	<p>O1. Supplier must provide license software for the complete XPS operation including option to make changes manually.</p> <p>O2. The system must have ability to be controlled remotely via internet/network.</p> <p>O3. There must be at least 5 numbers of data processing software user licences (or dongles) for XPS spectra analysis through peak identification using the library and peak fitting.</p> <p>O4. The latest version software must be provided with free update. The hardware should also be latest window-based system.</p> <p>O5. XPS polymer database/library must be provided.</p>
Calibration standards	Standard samples (Ag, Au, Cu and PET) must be provided to check the system calibration, resolution, sensitivity, etc.
Consumables and accessories	The system should be supplied with all the accessories and consumables to run for at least 5 years. More precisely, another two sets of Al monoanode assembly, filament assembly, charge neutralizer assembly, channel plate kit, ion gun grid filament assembly, filament assembly, must be provided as and when life time of existing one is over or in the event of their failure. Also complete set of backing kit for UHV, safety devices etc, should be provided.
Warranty	<p>Comprehensive Warranty for THREE years for entire system including all the attachments (such as UPS, chiller and compressor or any other third party component) of the system from the date of installation and commissioning. The breakdown period should not be counted as warranty period and must be extended after the expiry of initial warranty. The warranty must cover unlimited breakdown service calls and calibration.</p> <p>Service response must be within 48 h.</p> <p>The cost for additional TWO years of warranty and/or AMC after three years of warranty should be mentioned.</p>
Manuals	A complete set of instruction and service manual both in hard and soft copies must be provided along with the system.
Installation and training	After installation by the engineers of supplier, the system performance must be demonstrated to the customer's satisfaction as per the

	<p>specification detailed here.</p> <p>The supplier must provide basic training to the users for at least seven days and advanced training at least for five days after completion of installation.</p> <p>The supplier must be willing to extend a 15 days training to the technical officer who would look after the equipment at the place of manufacturer.</p> <p>A follow-up training must be provided by the supplier's engineer after six to nine months of installation.</p>
Pre-installation Requirements	Pre-installation requirements such as room size, EM tolerance limits, and vibration limits should be mentioned and verified by the supplier at the installation site as soon as PO is issued.
Additional Terms	<ul style="list-style-type: none"> <li>- The firm has to guarantee technical support for the entire system and supply of spares for a minimum period of 10 years from the date of installation.</li> <li>- Service response time must be &lt;48 hours.</li> <li>- Price break up for individual items should be provided in the price-bid. The prices of all optional items should be quoted separately.</li> <li>- Provision for on-line remote diagnosis of faults.</li> <li>- The firm must have atleast <b>two</b> installations of similar equipment within India to establish desired competence for maintenance.</li> <li>- List of users with copies of installation report and contact details (address and phone number) of the users, where instrument has been supplied in last three years should be provided.</li> <li>- Date of manufacturing of the equipment should be after the placement of order.</li> <li>- Original documents from the manufacturer supporting the specifications should be provided.</li> <li>- A detailed compliance statement indicating point-wise response of the vendor to every item in the specifications, optional items as well as terms should be provided.</li> <li>- Price-bid should have every item mentioned in technical bid.</li> </ul>

#### Optional Items to be quoted

Sl. No.	Optional Items
Optional item: 1	Air-sensitive sample transporter.
Optional item: 2	The supplier may provide a highly skilled full time person with suitable expertise for providing technical assistance and routine maintenance of the proposed XPS system attached with UPS for a period of 3-years from the date of installation of the system in the institute. The institute shall have no responsibility for his/her service liabilities. The expenses for such

	service may be quoted separately.
--	-----------------------------------

## ANNEXURE – II

### GENERAL TERMS AND CONDITIONS

(1) **Last Date of Submission of Sealed Bids (both technical and price bids, separately): 04-01-2018 up to 11:30 a.m.**

**The Technical bids will be opened on 04-01-2018 at 3.30 p.m. in the Office of the Central Research Facility, I.I.T. Kharagpur, India.**

(2) **Payment Terms & Performance Guarantee:** 90% payment will be made on submission of shipping documents and balance 10% payment will be made on installation, commissioning and submission of PBG for performance period if the payment is made by LC, otherwise after receipt of stores by any other payment mode. The security deposit so retained may be refunded on submission of Bank Guarantee towards Performance Guarantee valid for **3 years throughout warranty period plus sixty days** drawn on any commercial bank.

No advance/mobilization support, is payable against supply of stores.

In the event of failure to deliver the stores beyond the specified date, liquidated damages @ 1% per month or part thereof in respect of the value of stores will be deducted, subject to a maximum of 5%; alternately the order will be cancelled and the undelivered stores purchased from elsewhere at the risk and expense of the vendor.

(3) **Warranty/Guarantee & On-site skill support:** Minimum **three years** onsite comprehensive OEM warranty from the date of successful installation and commissioning. The firm has to guarantee supply of spares for a minimum period of 10 years. The OEMs/Authorized Distributors and Dealers must attach certificate about their after sales and service facilities, escalation support for on-call service or station engineer etc. AMC for 3 to 5 years to be quoted separately.

(4) **Delivery of Stores:** The store items be supplied within 30 days from the date of receipt of the Purchase Order or from opening of LC whichever is later.

(5) **Tender Fee:** An amount of **INR 5,000** (Rupees five thousand only) as tender fee (non-refundable) is to be paid. The payment shall be made by Demand Draft from any Nationalized Bank /Commercial Bank and paid in favour of "Indian Institute of Technology Kharagpur", payable at Kharagpur. **Bids without Tender Fee will not be accepted. This should be enclosed separately in an envelope and stapled with the Technical Bid document superscribing Tender fee.**

(6) **Earnest Money Deposit (EMD):** An amount of Rs. 2,00,000/- in the form of Demand Draft drawn in favour of “**Indian Institute of Technology Kharagpur**”, payable at Kharagpur or Fixed Deposit Receipt or Banker’s Cheque to be obtained from the Bidders. **E.M.D. should be enclosed separately in an envelope and stapled with the Technical Bid document superscribing EMD.** The bid security (EMD) is normally to remain valid for a period of 45 days beyond the final bid validity period. **Any bid without EMD will summarily be rejected. No interest is payable on EMD.** EMD will be refunded to the unsuccessful bidder on finalization of the tender process. The EMD of vendor awarded with the contract to be treated as part of security deposit towards Performance Guarantee and will be refunded back on submission of performance security. No interest is payable on Security Deposit. Security Deposit shall be forfeited if the selected vendor after award of contract fails to execute the same.

(7) **Price:** Domestic tenders are to quote and accept their payment in Indian currency. Indian agents of foreign suppliers are to receive their agency commission in Indian currency. Cost of imported goods, which are directly imported against the contract, may quote in foreign currency (currencies) and will be accordingly in that currency and the portion of the allied work and services, which are to be undertaken in India (like installation & commissioning of equipment) are to be quoted and will be paid in Indian currency. The price shall be quoted with cost break-up indicating the applicable prices for each component (custom duty, import duty, landing and clearing charges and commission to Indian agents). Excise Duty Exemption/Custom Duty Exemption Certificate and Inter-state transit Way Bill (Form-50) will be issued only in favour of the bidder and address as mentioned in the Purchase Order.

(8) **Bid:** Technical Bid and Price Bid should be submitted in two separate sealed envelope quoting reference number on the top of the envelope. Tender Fee and EMD should be enclosed with the Technical Bid documents, in separate sealed envelopes, stapled with the packet containing Technical Bid documents. The OEMs may either bid directly or Country/Regional/State Distributors /Dealers/Vendors, to quote authorized with valid authorization certificate, capability to sale and service of the products.

(9) **Acceptance of Tender:** The Authority of IIT Kharagpur does not bind itself to accept the lowest priced bid and reserves the right to reject any or the entire tender bids received without assigning any reason thereof.

(10) **Extra Features:** If the bidder provides any other extra features on the Hardware or Software which are not mentioned in the tender product specifications, then that shall be highlighted in clear terms, with documentary evidence/literature.

(11) **Compliance List:** The proposal be properly indexed and a compliance list against the technical specifications should be provided.

(12) **Service:** Response to ensure quality of services, the deputed Engineer/Technical officer from the OEM/Vendor must be well trained and must be in the payroll of the OEM/Vendor.



(13) **Installation and Commissioning:** Free of cost at IIT, Kharagpur. The OEM must ensure timely installation of X-Ray Photoelectron Spectrometer attached with ultraviolet photoelectron spectrometer and accessories with necessary support to the indenters, as per details and lists to be made available by the Stores & Purchase Section or the indenting Departments/Centres/Schools.

(14) **Validity of licenses:** Software's licensing price or policy (if any) shall be clearly mentioned.

(a) All licenses should be perpetual

(b) All the accessories shall be from the same OEM

(15) The OEM should be an ISO-9000 and ISO-14001 certified company with due credits to energy conservation and green earth compliance.

**(16) Relevant documents of the OEM shall be enclosed, along with the Technical Bid. Any explanation on this account shall be supported with documentary evidence from the principals.**

(17) **Conditional Offer** will not be accepted.

(18) **Period of Validity:** Bids shall remain valid for acceptance for a period of 120 days from the date of opening of the price bid.

(19) The benefit of any downward price revision (revision on account of budget/financial policy, tax revision, EPZ etc.) is to be given to IIT Kharagpur by the selected OEM/vendor.

**(20) Past Performance of the Vendors will be judged at the time of Technical Evaluation.**

(21) **The Institute does not bind** itself to offer any explanation to those bidders whose technical bids have not been found acceptable by the Technical Evaluation Committee of the Institute.

(22) **Bidders should** enclose the following documents:

(i) Certificate of Registration / Trade License

(ii) Enlistment form (R1), complete in all respect

(iii) Attested copy of PAN card, Service Tax, CST/VAT registration papers

(iv) Banker's Solvency Certificate

(v) Audited statement of accounts and IT returns for the last three years

(vi) Authorized Distributors/Vendors must submit appropriate authorization certificate and letter from their OEMs, for participation in the said tender.

(vii) Name and address of past satisfactory supplies or minimum three clients to whom such items/stores have been supplied should be mentioned in the technical bid.

(viii) Copy of mandatory test reports, national testing/reliability and endurance test reports etc., certified or conducted at the manufacturing site, granted by the bureaus/quality control departments/national testing laboratories.

(ix) Copy of product literature, for which the prices have been quoted.

(x) A write up on service and maintenance capability, mitigation of risks or breakdown and replacement capability, with the escalation support matrix suggested for the Institute. Vendors must indicate their sales and support service centre in India and their plan to address issues about services, maintaining minimum service inventory etc.

(xi) Signed copy of the tender document, with company seal, agreeing to the terms & conditions and declaration should be provided.

(xii) DGS&D enlistment certificate needs to be attached (applicable only for Indian Agents)

(23) **All tenders are to be handed over in a sealed box in the office of the Central Research Facility, IIT Kharagpur. The bids (technical and price bids) once submitted shall be the property of the Institute and shall not be returned to the vendor in future.**

(24) **The person/officer signing the tender/bid documents should be delegated with an appropriate Power of Attorney (duly endorsed by a Notary Public) by the Chief Executive Office/Managing Director of the Company to sign such documents.**

(25) **Opening of Price Bids:** The Price Bid(s) of only those vendor(s) who are found technically qualified will be opened and the same will be opened before the technically qualified vendor(s). **The date for opening of price bids will be notified separately.**

(26) Tenderer or his/her authorized representative (with proper authorization letter for attending opening of technical bids and also for opening of price bids) may choose to be present at the time of opening of Technical Bids/Price Bids.

(27) **IMPORTANT**

1. Director may accept or reject any or all the bids in part or in full without assigning any reason and does not bind himself to accept the lowest bid. The Institute at its discretion may change the quantity/upgrade the criteria/drop any item or part thereof at any time before placing the Purchase Order.

2. A bid submitted with false information will not only be rejected but also the OEM/vendor will be debarred from participation in future tendering process.

3. The OEMs/Vendors need to submit a certificate during opening of technical bids that they are not currently debarred or blacklisted in IIT Kharagpur for any supplies, products or services, or at present in any national organization or educational institute/ university.

4. In case of any dispute, the decision of the Director of this Institute shall be final and binding on the bidders.

5. For any query pertaining to this bid document, correspondence be addressed to :  
End user

Central Research Facility,  
Indian Institute of Technology Kharagpur- 721302,  
West Bengal, India

[Ph:03222-282480, Fax:03222-282481]

6. In case the due date for opening tender happens to be a holiday, the same will be opened on the next working day. The timings will however remain unchanged. Please Note that the Institute remains closed on Saturday & Sunday.

Chairman, Central Research Facility  
For Director, IIT Kharagpur

**Annexure-III**

## **INSTRUCTIONS & SPECIAL CONDITIONS**

*(To be returned by Tenderer along with the Tender duly signed)*

1. **GENERAL:** Tenderer(s), who are Indian Agents of OEMs should furnish a clear declaration as follows: We declare that I am/we are Accredited Agents of the suppliers aboard. DGS&D enlistment certificate needs to be attached (applicable only for the Indian Agents)

### 2. **DEVIATION FROM SPECIFICATIONS:**

*It is in the interest of the tenderer to study the specifications in the tender schedule thoroughly before quoting so that, if the tenderer makes any deviations, the same are prominently brought out in the body of the tender. If you need to add any optional items to your system in order to meet our specifications, you are requested to quote for the total including the option required to suit our requirements. Otherwise, your tender will not be considered at all.*

### 3. **PRICE:**

*For imported items both F.C.A./F.O.B. & C.I.F. Kolkata price by Air/Sea to be quoted. The price should be without Customs Duty and Excise Duty since IIT Kharagpur is exempted from the payment of Customs & Excise Duties. The Institute, if required will provide certificates for Import Items: The order acknowledgement should be issued by Fax within 3 days, and the item should be airlifted/boarded in sea within 30 days from the date of receipt of the Letter of Credit/copy of DD, as the case may be.*

*(i) Agency Commission, if any will be paid to the Indian agents in Rupees on receipt of the equipment and after satisfactory installation. Agency Commission will not be paid in foreign currency under any circumstances.*

*(ii) The unit prices should be for the same unit as indicated in the Schedule to tender enquiry and not for any other unit.*

*(iii) Prices quoted should be for supply, installation and commissioning at I.I.T. Kharagpur.*

*(iv) Discount, if any, should be indicated separately / prominently.*

*(v) Offers should normally be on fixed and firm price basis. Any clause making price variation will not be acceptable.*

*(vi) Where tenderer intends to quote in Indian Rupees, for imported items, the price should be quoted "FOR Destination" basis. Offer should be on "High Sea Sales" basis. No sale tax can be charged. Customs Duty applicable against **Declaration in Terms of Government of India Notification No.51/96-Customs dated 23.07.96 will be paid. Vendor will arrange payment of customs duty during clearance and the same will be reimbursed at actual against proper documents. Way Bill will be issued on request.** Seller has to undertake the responsibilities of customs clearance. All necessary supporting documents will be provided by the IIT*

Kharagpur. An agreement can be made in this regard, if the seller wishes, before customs clearance.

#### 4. TRANSIT INSURANCE:

The Purchaser will not pay separately for Transit Insurance.

#### 5. PAYMENT:

(i) Payment for indigenous items and also imported items quoted in Indian Rupees will be made only after completion of supply, installation and commissioning of the Systems to the entire satisfaction of the Purchaser. No Advance Payment will be made for Indigenous Purchase.

(ii) Where items are imported and the payment is desired to be made to their foreign Principals directly, the same will be made against Irrevocable Letter of Credit on submission of shipping documents or by any other mode of payment (i.e. Demand Draft/Sight Draft/Wire Transfer) after receipt of stores.

#### 6. TENDERERS SHALL SUBMIT ALONG WITH THEIR TENDER:

(i) Complete address and tele links for contact persons of principals and Indian agent offices dealing with this purchase.

(ii) Name and full address of the OEM's Banker and their swift code.

(iii) Proof of having ISO 9001 or other equivalent certification given by appropriate authorities.

(iv) If payment is made through LC, then LC opening charges inside and outside India, who will have to bear to be clearly mentioned

(v) Country of origin is to be provided for each item.

**7. EARNEST MONEY DEPOSIT: tenderer must submit an EMD of INR 2,00,000/- (Rupees Two Lakhs only (REFUNDABLE) IN THE FORM OF BANK DRAFT/BANERS' CHEQUE IN FAVOUR OF "INDIAN INSTITUTE OF TECHNOLOGY, KHARAGPUR". It should be placed in the cover containing of Technical Bid.**

- 1) **8. TENDER FEE: TENDER FEE OF TOTAL INR 5000(Rupees five thousand only) [NON- REFUNDABLE] IN THE FORM OF BANK DRAFT MUST ACCOMPANY THE TENDER. It should be placed in the cover containing the Technical Bid.**

#### 9. PERFORMANCEBANK GUARANTEE:

On behalf of the Principal, Indian Agent must be able to provide Performance Bank Guarantee of the amount equivalent to the 10% of the cost of equipment from any Nationalized Bank. The Performance Bank Guarantee is required for the entire period of Warranty. Performance Bank Guarantee must remain valid for a period sixty days beyond the expiry of theWarranty Period. Bid security will be refunded back on submission of performance Security.

#### 10. GUARANTEE:

The tenderer has to declare that the goods sold to the buyer under this contract shall be of the best quality and workmanship and shall be strictly in accordance with the specifications.

**11. JURISDICTION:**

All questions, disputes, or differences arising under, out of or in connection with the contract, if concluded, shall be subject to the exclusive jurisdiction at the place from which the acceptance of Tender is issued i.e. Jurisdiction of KOLKATA HIGH COURT. Acceptance to this effect is also necessary at the time of opening of Technical Bid.

**12. ACKNOWLEDGMENT:**

It is hereby acknowledged that we have gone through all the points listed under "Instructions & Special Conditions" outlined above, and those in the accompanying note on "Important Conditions", and we agree to abide by them under the penalty of permanent disqualification for Tender participation and for related penal actions for non-abidance of the conditions.

13. Interested vendors must be able to supply adequate spares and consumable during **three year warranty**. Vendor should also ensure trouble free service and performance for another **two years beyond three years Comprehensive Warranty Period** with adequate spares and accessories.

SIGNATURE OF TENDERER ALONG WITH  
SEAL OF THE COMPANY WITH DATE

NOTE: ONE AGENT CANNOT REPRESENT TWO SUPPLIERS OR QUOTE ON THEIR BEHALF AT ONE TIME.

In case of import all the particulars required for opening Letter of Credit (L/C) should be furnished in clear terms. After opening of L/C, based on our tender, no more requests for amendment of L/C will be entertained. Total price including supply, installation and commissioning at I.I.T., Kharagpur, should be quoted. The cost of items that may be supplied from indigenous resources and also installation part if installation is done by Indian Agents should be stated in Indian Rupees.

1. In case of Import, bidders are requested to quote FCA/FOB price mentioning Freight & Insurance charges separately. Ex-Works pricing may be considered.
2. Warranty should be valid for Three years for all items including third party items.
3. Quotations should be submitted in Separate Sealed Cover (with Sealing Wax). The quotation should be in typed format. No manual entries are allowed and any manual entry will entail rejection at the opening stage itself.
4. Before submitting vendor should ensure that they have submitted the following in the technical bid, the documents as mentioned hereunder: Technical Literature, Specification documents, **Point-by-point Compliance Statement**, General Terms & Conditions, Acceptance of Kolkata High Court Jurisdiction, Name and address of 5 organizations of which 1 must be in India where the OEM has installed 1 such equipment in the last 2 years and a copy of Price bid with prices masked.

**The Commercial bid shall have all the commercial documents, agreements and Price bid. Prices should be given component wise to enable the Institute to decide the final configuration.**



**TECHNICAL BID DOCUMENT**

**FORMAT TO BE FILLED BY THE MANUFACTURERS/ INDIAN AGENTS (ON BEHALF OF THEIR FOREIGN PRINCIPLES) OR THEIR AUTHORIZED COUNTRY DISTRIBUTORS/ REPRESENTING DEALERS/ REGIONAL AGENCIES, SUBMITTING TENDER FOR SUPPLY OF X-ray PHOTOELECTRON SPECTROMETER (XPS) ATTACHED WITH ULTRAVIOLET PHOTOELECTRON SPECTROMETER (UPS) FOR IIT KHARAGPUR (CENTRAL RESEARCH FACILITY)**

1. Name of the Tenderer :
2. Status of the Tenderer: (attach documents, if registered company/partnership/proprietaryship)
3. Whether OEM/representing foreign principle :  
(attach copy of certificate/authorization)
4. Details of key top official/authorized official:  
(attach details)
5. Details of tie-ups for supply/services, if any:  
(Attach details, agreements, escalation matrix)
6. Income Tax and Service Tax returns of previous three assessment year (copy):
7. Financial status of bidder and/or his associates including Annual Report & Balance Sheet/ Statement of Account of past three years with Registration of Companies (ROC) receipts duly authenticated by Chartered Accountant :
8. Current list/address of clients where similar material has been supplied and successfully working:
9. Name of the vendor's three largest clients, to whom similar products and services were extended & amount of transaction/annual bills to such clients:
10. Name and address of Vendor's bankers and attach a Solvency Certificate from the Bank for a minimum amount of Rs. 20.00 Lakh:
11. Income Tax Permanent A/c No. (attach copy) :
12. **Details of EMD/Bank Draft No., issuing branch and date:**  
Certified that all above information are correct to the best of my/our information, knowledge and belief.

-----Dated  
signature & seal of the Authorized person of OEM/Vendor

**NOTE: This is to be submitted in a separate sealed envelope** super scribing "TECHNICAL BID", Notice inviting Tender No.IIT/Dept/ **Name of Item**/2014-15 and name of the bidder. All technical documents like literature, catalogues etc., are to be put in the same envelope. Price bid of that company/firm only will be opened which

do technically qualify, for further consideration. **Attach all relevant documents in the same serial order as above, properly indexed, duly signed and sealed.**



**DECLARATION**

1. I, -----Son /Daughter of Shri -----  
-----  
Proprietor/Partner/CEO/MD/Director/  
Authorized Signatory of M/s. ----- am competent  
to sign this declaration and execute this tender document.
2. I have carefully read and understood all the terms and conditions of the tender  
and hereby convey my acceptance of the same.
3. The information/ documents furnished along with the above application are true  
and authentic to the best of my knowledge and belief.
4. I/we am/are well aware of the fact that furnishing of any false information/  
fabricated document would lead to rejection of my tender at any stage besides  
liabilities towards prosecution under appropriate law.
5. Each page of the tender document and papers submitted by my Company is  
authenticated, sealed and signed, and I take full responsibility for the entire  
documents submitted.

----- Signature of the Authorized Person

Date: ----- Full Name: -----

Place: ----- Company Seal: -----