

# Corrigendum

**Web-tender No: IIT/SRIC/EE / IHL / SM / 2017-18/EQ-7, Dated: 27.12.2017**

- Four AC/DC current clamps with following specification should be added optionally.

Current Clamp	
Dc range	0.1 – 100A
AC range	0.1A-70A
Bandwidth	100kHz
Maximum conductor diameter	11.8 mm
Accuracy	0.1A to 10A; dc to 1 kHz: $\pm 3\% + 50\text{mA}$ , up to 100A add 15%; up to 20kHz add 12%

- The following portion should be excluded from the given specification.

**Important Note:**

The instrument should be upgradable in nature. The following hardware and software upgrade options must be available for upgrade at a later date.

1. Bandwidth Upgrade,
2. 16 Digital Channel Logic Analyser,
3. Spectrum Analyser,
4. Arbitrary Function Generator,
5. Protocol Analyser,
6. Limit/Mask testing application,
7. Power Analysis application, etc