

P.K.Sinha Center for Bioenergy
Indian Institute of Technology
Kharagpur – 721 302, West Bengal, India

Corrigendum

Tender Reference: IIT/SRIC/PKSC/EFC/RB/2017-18/EQ/1 Dated: 23/04/18 the following changes may be noted:

- 1) Horizontal Autoclave with drive.
- 2) Electrical Power: Sufficient wattage for bringing the temperature up to 121-130 °C inside the chamber by electrical/ boiler/ any other means to generate steam within a reasonable period of time that may vary from 30 min to 1 h.

All other terms and conditions of above mentioned tenders will remain unaltered.



Rintu Banerjee
Prof. Rintu Banerjee
Principal Investigator
27/4/18